

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAY 11 2010

Supervisor/Operator Performing the Verification Procedure:

DataMaster cdm S/N 130187 ✓

Name SGT. JOHN HAMILTON ✓

ID# 3192 ✓ Date 5/14/10 5/13/10 6/23/10

A

Agency PETERSBURG POLICE DEPARTMENT

Phone # 772-3838

Instrument Location PETERSBURG

B

Alco S/N 78820 ✓

Target Value .083

High Pressure 800 PSI

Alco Test Values

.086

1<sup>st</sup> Alco

.086

2<sup>nd</sup> Alco

Signature

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 2nd day of July, 2010.

(Notary Seal Stamp)

Carolyn M. Noland

Notary Public, State of Alaska

Commission Expires with Office

Target value not programmed. Breath  
Personnel needed to authenticate VOC  
for court. BT9 06/08  
BMB 6/23/10



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130184

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130184

MAY 04, 2010

OPERATOR'S NAME:  
HAMILTON JOHN A  
OPERATOR'S NUMBER: 3192  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/NI :

A  
O.L. #: A  
DEPT/AGENCY: PS61  
CASE/REPORT: A  
TEST TYPE: U  
ALCO TARGET VALUE: .003  
ALCO S/N: 78820

### — BREATH ANALYSIS —

.001 ADJUSTED FOR 30.33 in  
ALCO TARGET .002 06:38  
BLANK TEST .000 06:39  
INTERNAL STANDARD VERIFIED 06:39  
ALCO TO 30.33 in .006 06:40  
BLANK TEST .000 06:41  
SUBJECT SAMPLE .000 06:41  
BLANK TEST .000 06:42  
ALCO TO 30.33 in .006 06:42  
BLANK TEST .000 06:43

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130184

MAY 04, 2010  
TIME 06:44

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/28/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 40c  
BAROMETER: 30.35 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=?@ABCDEF6  
HIJKLMNPOQRSTUVWXYZ[\]^\_`abcdefghijklmnopqrstuvwxyz{|}~